

# An Introduction of Myo Armband and Its Comparison with Motion Capture Systems

Junghun Cho<sup>1</sup>, Jang Hyung Lee<sup>1</sup>, Kwang Gi Kim<sup>1\*</sup>

## Abstract

Recently, ways for accurately measuring the three dimensional movements of hand are actively researched so as to utilize the measurement data for therapeutic and rehabilitation programs. This research paper aims to introduce a product called Myo Armband, a wearable device comprised of a 3-axis accelerometer, a 3 axis gyroscope, and electromyographic sensors. We compare Armband's performance with that of the Motion Capture System, which is known as a device for providing fairly accurate measurements for angular movements of objects. Dart throwing and wrist winding motions comprised movement scenarios. This paper also discusses one of Armband's advantages - portability, and suggests its potential as a substitute for previously used devices. Decent levels of measurement accuracy were obtained which were comparable to that of three dimensional measurement device.

**Keywords:** Myo Armband, Gyroscope, Accelerometer, Motion track system, Wearable device.

## I. INTRODUCTION

The three-dimensional movement of hand with respect to wrist joint can be classified into one of four categories by assumed direction – flexion and extension (FE), and radial and ulnar deviation (RUD). Among the various joints of the body, wrist joint is one of the most complex structures, with its 15 bones including the distal ends of the radius and ulna. But measurement of the extremes of flexion-extension and radioulnar deviation of the wrist joint provides a limited picture of the static ability of the joint to move in the orthogonal planes. It also is difficult to ascertain the relationship between function of a complex joint like the wrist and these 2-dimensional measurements [1], [2].

Currently, a most common way of measuring wrist's full range of motion includes goniometer. Data is read directly from the device, which is adjusted accordingly on the subject's body. The advantages of this method include maximal portability and high accessibility due to its small-scale and light weight. It is the least expensive option for measuring joint movements, with cost in the range from about 5 USD to 100 USD [3]. Recently, smartphone goniometry applications have become available to clinicians, with error range of only about 1.7 degrees which is comparable to that of data gathered from prevalent goniometry [4]. A disadvantage of goniometer is its rather low level of accuracy. Since the accuracy of goniometry

measurement is dependent on a host of factors such as variations among different subjects, methods of application, differences among the kinds of motions measured, and individual deviations by clinicians gathering the data, it is often hard to meet expected level of accuracy using goniometer [5]. The center of the bone is the ideal point to place devices on for measuring FE and RUD motions but goniometer alone, it is impossible to pinpoint the ideal point of measurement. Measuring is performed with the device on the surface of the arm, which is constantly affected by several factors including muscle and skin contractions. This is also the reason why recently-released digital type goniometers are not considered accurate either. Another disadvantage of goniometer is that real-time tracking of motion is impossible. This discourages its use in the fields such as robotics or virtual-reality (VR).

The reason the tool is widely used even with such a high error range is because a difference of  $\pm 10$  to 15 degrees in range of motion (ROM) over time does not signify a meaningful change when it comes to comparatively examinations of patients [5].

Among the various apparatus for measuring ROM, another well-known one that overcomes these flaws is the three-dimensional (3D) motion capture system. High accuracy of this system has been proved by several previously released studies [6]. A 3D motion capture device, specifically Optitrack, coupled with Micron Series

---

**Manuscript received April 02, 2018 ; Revised May 23, 2018 ; Accepted May 26, 2018. (ID No. JMIS-2018-0026)**

Corresponding Author (\*): Kwang Gi Kim, Dept. Biomedical Engineering, School of Medicine, Gachon University, Incheon, 21565, Republic of Korea, [kimkg@gachon.ac.kr](mailto:kimkg@gachon.ac.kr)

<sup>1</sup>Dept. Biomedical Engineering, School of Medicine, Gachon University, Incheon, 21565, Korea, [charlie.cho2000@gmail.com](mailto:charlie.cho2000@gmail.com), [leejh161@naver.com](mailto:leejh161@naver.com)

calibration tools, can produce fairly precise measurements down to sub-20 $\mu$ m tolerance. It may be possible to set the measuring points at the centers of the bones by a post-experiment calibration procedure. Yet, it requires ample space, costly devices and special programs.

Myo Armband (Thalmic Labs, Inc) is a wearable device that consists of an inertial measurement unit (IMU) and eight electromyographic (EMG) sensors. IMU in turn consists of a 3-axis gyroscope, a 3-axis accelerometer and a magnetometer. They together measure muscle tension and arm motion. Data communication is done wirelessly through Bluetooth 4.0 connector. With a light weight of 93 grams and adjustable circumference ranging from 19 centimeters to 34 centimeters, it was designed in consideration of wearer's comfortability. The aim of this study was to compare the accuracies of three dimensional movements recorded with Myo Armband with those of 3D motion capture system.

## II. METHOD

Traditionally, kinematic studies of the wrist have been focused on the orthogonal flexion/extension and radial/ulnar deviation motions. During wrist flexion/extension, the scaphoid and lunate flex and extend with the wrist. During wrist radial/ulnar deviation, both bones continue to rotate in the flexion/extension plane; flexing during wrist radial deviation and extending during wrist ulnar deviation

In the human wrist, dart throwing and circumduction gesture provides a more accurate representation of the capabilities of wrist movement [7], [8] because it combines flexion, extension, and radioulnar deviation into a circular motion without simultaneous supination or pronation of the forearm. These coupling of motions are vital in activities of daily living. These activities involve a physiological axis of wrist motion [9], [10] in an oblique plane [11], [12] rather than in the orthogonal flexion-extension or radioulnar deviation axis.

Since we wanted to observe how accurately Myo Armband measures hand-wrist movements, each participant was instructed to make 10 dart throwing motions and 10 circumduction motions. Before each gestures, all participants went through demonstration and was rectified, if needed. Same gestures were done for both Optitrack and Myo Armband.

### 2.1. Optitrack

The use of low cost optical motion capture (OMC) multi-camera systems is spreading in the fields of biomechanics research [13].

Motion capture systems have been used extensively in gait analysis, providing accuracy of within 1% error for 120 mm translations [14]. Motion capture cameras have light emitting diodes that emit light into a capture area. Markers in the capture area reflect the light back to the camera which then determines the size and location of the marker [34]. Using multiple markers and cameras, the motion capture system is then able to triangulate the actual position of markers in 3D space.

In this experiment, OptiTrack's Flex 13 (Leyard Company Inc.) was used. This model, Flex 13, has been shown to provide clinically accurate and reliable results in previously studied cases, therefore, was employed. The cameras were of 1.3 Megapixel resolution (1280 x 1024 pixel) and were equipped with stock lenses (5.5 mm focal length, 56° horizontal and 46° vertical field of view). For the gathering and the processing of motion tracking data, Motive from the manufacturer (Tracker v. 1.10.0, NaturalPoint, United States) was used.



Fig. 1. Optitrack's Flex 13 model. In this test, it was used to measuring joint movement angles

To foreclose any possibility of infrared (IR) interference caused by the sunlight, the experimental setup was employed in a secluded room located at the basement, that of Gil Hospital, Incheon, Korea. Six motion capturing cameras were placed 1.8 meters above the ground, in such a way that they would form a circle of approximately 2.5 meters' diameter (see Figure 2). The frame rate was set at 120 Hz. The built-in calibration wand CW-500 was used to calibrate the system. its nominal size is 500 mm. The true size of the calibration wand was also measured with an ATOS II Triple Scan MV320 (GOM, Braunschweig, Germany). The wandng process involved moving the calibration wand throughout the entire capture volume. At the end of calibration, the number of samples exceeded 20,000 frames for each camera.




Fig. 2. Six cameras were set up 1.8 meters above the ground forming a 2.5m-diameter circle (above). Then, the test environment was demonstrated on Motive

Three retroreflective markers were attached midway between the distal radius at the level of the ridge between the radioscaphoid fossa and the radiolunate fossa [16]. Position and rotation values of the center of the rigid body were exported as XYZ translational coordinates in comma-separated values (csv) format, and later processed by using Microsoft's Excel (Microsoft, ver. 2013) and Origin Lab (Origin Lab, ver. 8) for the visualization of the hand movements.




Fig. 3. Three retroreflective markers were attached to a 3D-printed sculpture (left). Then, their center point was calculated (right).

## 2.2. Myo Armband

As mentioned previously in introduction, Myo Armband is a wearable device that consists of an inertial measurement unit (IMU) and eight electromyographic (EMG) sensors. Multi-accelerometer systems have already

shown the ability to recognize activities and detect falls with high accuracy.



Fig. 4. Myo Armband (Thalmic Labs Inc) was held in a way that its sensor would be placed on the same location as those of retroreflective markers in Optitrack experiment.

To minimize the errors due to the inconsistency of constant, participants were asked to hold the armband in a way that the accelerometer and gyroscope inside the armband would be placed between the radioscaphoid fossa and the radiolunate fossa. Then, they were asked perform 10 dart throwing and 10 circumduction gesture.

Myo Armband's data was recorded using a program named 'Myo Data Capture Windows', which was written and published by the developers of the Myo Armband, Thalmic Lab. The program detects quaternion data from Myo Armband then transforms it into Euler Angles by following formula:

$$\text{Roll} = \phi = \tan^{-1} \left[ \frac{2(q_0q_1 + q_2q_3)}{1 - 2(q_1^2 + q_2^2)} \right]$$

$$\text{Pitch} = \theta = \sin^{-1} [2(q_0q_2 - q_3q_1)]$$

$$\text{Yaw} = \varphi = \tan^{-1} \left[ \frac{2(q_0q_3 + q_1 + q_2)}{1 - 2(q_2^2 + q_3^2)} \right]$$

## III. RESULTS

Through this experiment, we were able to visualize and compare hand and arm's FE and RUD movements measured respectively via Myo Armband and Motion Capture Systems. The movements were captured from three directions: front, side and top. The data captured through Myo Armband is to be compared to data captured with Optic track, which we consider as an ideal value.

#### IV. LIMITATIONS

In this experiment, there were two obstacles that distracted us from getting ideal results. These are: (1) variabilities in factors such as light reflection, camera setup, and capability of experimental subjects during the motion capturing procedure; and (2) the contraction of arm muscles. Although 3D motion capturing is generally known as a highly-accurate object tracking method, it should be noted that certain conditions are to be met to achieve ideal results. There are various types of 3D motion capturing system, but in this experiment, we adopted ‘optical system’ that makes use of markers coated with retroreflective material. This system requires secluded, ample space with no extra light reflecting object, a plenty number of fixed cameras and proficiency of measurer.

To minimize error, the test was done in a controlled room solely occupied by measurer at the Hospital room in the basement. The cameras were set up in a circular shape with a circumference of about 2 meters so that the trajectory of retroreflective markers could be captured from all directions.

Despite these efforts, there were few frames in which one or more markers were not recognized. Synthetic frames were supplemented by the program Motive’s built in function. Although errors stemming from the missing frames do not significantly affect the overall results, it may still be reduced by averaging over the numbers from motion capturing cameras during the test procedure.

Another factor that may affect the test results is the repeatedly contracting muscles of the subject’s arm. Since the test procedure involves drastic movements of arm, markers are susceptible to displacements along with the contracting skin, therefore reducing accuracy of the test. To reduce the resulting errors, we created a sculpture using 3D-Printing, where we attached the markers so there won’t be direct contact between the skin and the markers.

#### V. CONCLUSION

The incorporation of elements including accelerometer, gyroscope, and electromyographic (EMG) sensors makes Myo Armband a versatile device deployable in a range of applications. Starting from simply as a substitute for computer mouse, it may also be used as a maneuvering device for game or as a drone controller. The wireless data communication through Bluetooth is also a huge advantage, since it grants users spatial freedom.

Myo Armband’s accuracy has been directly and indirectly verified in several studies. For instance, researchers at Johns Hopkins University is now inventing a mind controlled robotic arm that can touch and feel, and




Fig. 5. The front, side, and top view of dart throwing movements recorded with Myo Armband (left) and Optitrack (right).




Fig. 6. The front, side, and top view of circumduction movements recorded through Myo Armband (left) and Optitrack (right).

they are using two Myo Armbands to measure the electromyograph (EMG) of muscles to see the corelationship between EMG and brainwaves [17]. Considering the level of accuracy required in this kind of experiment, the researchers' selection for Myo Armband testifies its performance.

In another study, Myo Armband was used to collect the body signals of people suffering from Parkison's syndrome [18]. Parkison's syndrome is a disease caused by the loss of nerve cells in specific parts of the brain. They put Myo Armband on patients' body to record and analyze the EMG data.

The company ADORA used Myo Armband as an assist device during surgery. During medical operation procedures, the range of surgeon's arm movement is rather limited, so the assistance of staff members is needed. Myo Armband, if worn inside the operating gown, may be used as a control device for medical equipment.

Previously employed sensors were expensive, and required a lot of curation efforts to achieve an acceptable level of data accuracy. Myo Armband, with its fair price and decent performance, is expected to find increased uses in various fields of science, especially in medical engineering.

#### REFERENCES

- [1] Harvinder P. Singh, Joseph J. Dias, Harm Slijper, Steven Hovius, "Assessment of Velocity, Range, and Smoothness of Wrist Circumduction Using Flexible Electrogoniometry," *The Journal of Hand Surgery*, Vol. 37, No. 11, pp. 2331-2339, 2012.
- [2] Tomaino MM, Miller RJ, Burton RI. "Outcome assessment following limited wrist fusion: objective wrist scoring versus patient satisfaction" *Contemp Orthop*, 28, pp.403-410, 1994.
- [3] C.C. Norkin, D.J. White, "Measurement of Joint Motion: A Guide to Goniometry," F.A. Davis Company, 2009.
- [4] R.A. Dos Santos, V. Derhon, M. Brandalize, D. Brandalize, L.P. Rossi, "Evaluation of knee range of motion: Correlation between measurements using a universal goniometer and a smartphone goniometric application", *J Bodyw Mov Ther.*, Vol.21, No.3, pp. 699-703, 2017.
- [5] R.L. GAJDOSIK, R.W. BOHANNON, "Clinical Measurement of Range of Motion Review of Goniometry Emphasizing Reliability and Validity," *Phys Ther.* Vol.67, No.12, pp.1867-1872, 1987.
- [6] M. Yazdifara, M.Reza, Y. Jamaluddin, M.I. Esata, M. Chizaria, "Evaluating the Hip Range of Motion Using the Goniometer and Video Tracking Methods," *Procedia Engineering*, Vol. 68, pp. 77-82. 2013.
- [7] Ojima H, Miyake S, Kumashiro M, Togami H, Suzuki K. "Ranges of dynamic motion of the wrist in healthy young and middle-aged men," *Ergonomics*, Vol. 35, pp.1467-1477, 1992.
- [8] RAWES, M. L., RICHARDSON, J. B., DIAS, J. J. "A new technique for the assesof wrist movement using biaxial fleaxible electrogoniometer," *J. Hand Surg.*, 21, pp. 600 603, 1996.
- [9] Taleisnik J. "The wrist. New York: Churchill Livingstone," pp. 41-42, 1985.
- [10] Wolfe SW, Crisco JJ, Orr CM, Marzke MW. "The dart-throwing motion of the wrist: is it unique to humans?," *J Hand Surg*, Vol.31, pp.1429 -1437, 2006.
- [11] Moritomo H, Apergis EP, Herzberg G, Werner FW, Wolfe SW, Garcia-Elias M. 2007 IFSSH "Committee report of wrist biomechanics committee: biomechanics of the so-called dart-throwing motion of the wrist," *J Hand Surg*, Vol.32, pp.1447-1453, 2007.
- [12] Palmer AK, Werner FW, Murphy D, Glisson R. "Functional wrist motion: a biomechanical study," *J Hand Surg*, Vol.10, pp.39-46, 1985.
- [13] T. Bihl "Planning of spatially-oriented locomotion following focal brain damage in humans: A pilot study," *Behav. Brain Res.*, Vol.301, pp. 33-42, 2016.
- [14] Thewlis D, Bishop C, Daniell N, Paul G. "Next generation low-cost motion capture systems can provide comparable spatial accuracy to high-end systems," *J Appl Biomech*, Vol.29, pp.112-117, 2013.
- [15] Richards JG. "The measurement of human motion: a comparison of com- mercially available systems" *Hum Mov Sci*, Vol.18, pp.589-602, 1999.
- [16] W. Ge, ISB. "Recommendation on definitions of joint coordinate systems of joint coordinate systems of various joints for the reporting of human joint motion- Part II: shoulder, elbow, wrist and hand," *Journal of Biomechanics*, 2004.
- [17] G. Hotson, R.J. Smith, A.G. Rouse, M.H. Schieber, N.V. Thakor, B.A. Wester, "High Precision Neural Decoding of Complex Movement Trajectories Using Recursive Bayesian Estimation with Dynamic Movement Primitives", *IEEE Robot Autom Lett*, 1(2):676-683 (2016).
- [18] S. Spasojevic, T.V. Ilic, I. Stojkovic, V. Potkonjak, A. Rodic, J. Santos-Victor, "Quantitative Assessment of

the Arm/Hand Movements in Parkinson's Disease Using a Wireless Armband Device", Front Neurol. Vol.11; No.8, 2017.

#### Authors



**JungHun Cho** was a student of Holy Angel University and his major was Aeronautical Engineering. He was a student researcher at Biomedical Research Center of Gil Hospital, Gachon University. His research interests include biomedical devices, artificial intelligence, and cognitive science.



**Jang Hyung Lee** works as a researcher at Dept. of Biomedical engineering of Gachon University and at Gil Hospital. He graduated from Arizona State University in 2011 with a major in computer science. Prior to joining Gil Hospital, He was a research staff member at S1 corp, S. Korea. He's been researching on application of artificial intelligence methods towards various domains including biometrics and biomedical sciences.



**KwangGi Kim** works now for Gachon University of Medicine as professor of Gachon University, Dept. Biomedical Engineering. He graduated Seoul National University in 2005. He had finished PostDoc. in Wasington University in St. louis in 2003-2005., he has been investigated on Computer aided diagnosis and Artificial intelligence as a researcher at National Cancer Center (2007-2017), he has studied on medcial robot and Artificial intelligence medcial imaging .